




Datasheet - characterization

Trace Code*: **BioPixS0080**

Model	OCTML01
Layer Thickness	50 μm
Number of layers	>10
Substrate	BioPixS Phantom
Substrate optical properties	Absorption = 0.19 cm^{-1} , Reduced scattering = 8 cm^{-1} @780 nm
Phantom dimension	22 x 22 x 5 mm
Layer dimension	7 mm (diameter), 0.5 mm (height)
Batch code	OCTML01-P-HCD06
Layer optical properties	Absorption = 0.05 cm^{-1} , Reduced scattering = 3 cm^{-1} @780 nm
Refractive index	1.5
Refractive index of transparent encapsulation	1.41
Surface finish^	Polished
Application	Depth resolution, signal contrast vs depth test



Batch test image:



OCT test system specifications: Spectral Domain OCT Imaging System, 1300 nm, axial resolution is 5.5 microns, lateral resolution of 13 microns.

*Cite/Reference Trace code in your publications to make your research traceable. This will help the peer community to trace the phantoms used in your research.

BioPixS Limited

3 Monastery Mews, Blarney Street
Gurrabraher T23 HW11 Cork, Ireland

<https://biopixstandards.com/>
VAT Reg. No: 3689183LH

Datasheet - characterization

^for matte finish visit <https://biopixstandards.com/products/oct-test-targets-and-phantoms/>

Phantom image:

